

GHANA

INTENSIVE CARE WARD: AMENDMENT OF NUMBER OF SETS OF RESUSCITATION / MONITORING EQUIPMENT

1. ISSUE PAPER THEME: Medical

2. SUMMARY / BACKGROUND

The COE Manual(A/75/121) Chapter 3, annex C, Appendix 5.1, page 92 Intensive Care Ward provides Resuscitation/Monitoring equipment. Amendment of number of sets of resuscitation/monitoring equipment from one (1) to two (2) in the intensive care ward.

3. DETAILED PROPOSAL

The COE manual, Chapter 3, Page 92 Annex C, appendix 5.1 B, the number of ICU beds required for Level II hospital is two (2). However, the number of resuscitation equipment required is 1 set. Amendment of number of sets of resuscitation/monitoring equipment from one (1) to two (2) in the intensive care ward is important because it is necessary for each bed to have its own set of monitoring/resuscitation equipment. There is a GFMV implication of \$37,106.00 if this proposal is adopted.

4. PROPOSED MANUAL TEXT

1. Chapter 3, annex C, appendix 5.1, Pg. 92

<i>Facility</i>	<i>GFMV</i>	<i>Item</i>	<i>Quantity</i>	<i>GFMV</i>
B. Intensive care ward	88,089	iii. Resuscitation/monitoring equipment	2 sets	
		Trolley with drugs		4 378
		Defibrillator		17 514
		Ventilator		14 228
		Intubation sets		3 284
		Infusion pump		9 850
		Suction pump		2 190
		Multiline vital signs monitor		21 892
		Oxygen cylinders		876